

English

30 x .35g Moisture activated temporary filling (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Remove cap.
- 2) Place tip into C-R® Syringe.
- 3) Squeeze syringe with a slow, steady pressure and inject into a moist cavity.
- 4) Tamp down with a wet instrument and dismiss patient. On return visit, Tempit® may be removed with an explorer.

Deutsch

30 x 0,35g Provisorisches Verschlussmittel, selbsthärtend (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Kappe des C-R® Tip entfernen.
- 2) C-R® Tip in eine Centrix Dosierpistole einsetzen.
- 3) Material mit leichtem gleichmäßigem Druck ausdrücken und in die feuchte Kavität applizieren.
- 4) Okklusalfäche mit einem feuchten instrument konturieren. Tempit® läßt sich während der endgültigen Sitzung leicht mit einer Sonde entfernen.

Français

30 x 0,35g Plombage provisoire à activation humide (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Retirez le capuchon.
- 2) Placez l'embout dans la seringue C-R®.
- 3) Appuyez sur la seringue et injectez lentement dans la cavité humide en maintenant une pression constante.
- 4) Tassez à l'aide d'un instrument mouillé et laissez partir le patient. A la visite suivante du patient, vous pouvez retirer Tempit® à l'aide d'un outil d'exploration.

Español

30 X 0,35g Empaste temporal activado con humedad (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Quite la tapa.
- 2) Coloque la punta en la jeringa C-R®.
- 3) Presione la jeringa con una presión lenta y constante e inyecte dentro de una caries húmeda.
- 4) Apriete el material hacia dentro con un instrumento mojado y deje ir al paciente. En la próxima visita, Tempit® puede ser quitado con un explorador.

Italiano

30 x 0,35g Otturazione temporanea attivata ad umido (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Rimuovere il tappo.
- 2) Collocare la punta sulla siringa C-R®.
- 3) Premere lentamente lo stantuffo della siringa esercitando una pressione uniforme ed iniettare il preparato nella cavità umida.
- 4) Pigiare il preparato con uno strumento umido e concludere la visita. In occasione dell'appuntamento successivo, Tempit® può essere rimosso con uno specillo.

Nederlands

30 x 0,35g Door vocht geactiveerde tijdelijke vulling (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Verwijder het dopje.
- 2) Plaats de punt in de C-R® injectiespuit.
- 3) Druk langzaam en met gestadige druk op de injectiespuit en injecteer in het vochtige gaatje.
- 4) Aandrukken met een nat instrument en de patiënt naar huis sturen. Bij de volgende visite kan Tempit® verwijderd worden met een onderzoekingsinstrument.

Dansk

30 x 0,35g Midlertidig fugtaktivert tandfyldning (REF 310060)

10°C - 24°C 50° - 75° F

- 1) Tag hættten af.
- 2) Sæt spidsen i C-R®-sprøjten.
- 3) Tryk på stemplet med en langsom, jævn bevægelse, og sprøjt præparatet i den fugtige kavitet.
- 4) Stamp til slut på fyldningen med et fugtigt instrument.
Tempit® fjernes med en skarp sonde ved patientens næste besøg.

Svenska

30 x 0,35g Provisorisk fuktaktiverad plomb (REF 310060)

10°C - 24°C 50°F - 75°F

- 1) Avlägsna hatten.
- 2) För in spetsen i en C-R®-spruta.
- 3) Injicera plomben i den fuktade kaviteten genom att trycka ihop sprutan i en jämn och långsam rörelse.
- 4) Packa till plomben med ett fuktigt instrument för att avsluta behandlingen. Tempit® avlägsnas med hjälp av sond vid återbesöket.

SECTION I - PRODUCT / COMPANY IDENTIFICATION

Product Name: TEMPIT® TEMPORARY FILLING MATERIAL
Synonyms: Temporary filling material
Part (Item) Number: 310060
Company Name: Centrix Inc.
Address: 770 River Road
City, State, Zip: Shelton CT, 06484
Emergency Telephone Number: 203-929-5582
Information Telephone Number: 800-235-5862
Date Revised: 5/2006

SECTION II - COMPOSITION / INFORMATION ON INGREDIENTS

Hazardous Components:	%	CAS No.	SYMBOL	R	EINECS No.
D.M.T.G (Glycol)	<12%	143-24-8			205-594-7
Ethyl Methacrylate Polymer	<3%				
Calcium Sulphate	<59%	10034-76-1			231-900-3
Zinc Oxide	<11%	1314-13-2			
Barium Sulphate	<17%	7727-43-7			
Silica (Amorphous)	<2%	7631-86-9			231-545-4

SECTION III - HAZARDS IDENTIFICATION

For potential health hazards: See section 11 below.
No other significant hazards to man or the environment are reasonably foreseen.

SECTION IV - FIRST AID MEASURES

Skin: Wash with soap and water.
Eyes: Flush with copious amounts of water. If irritation persists, seek medical attention.
Ingestion: N/A
Inhalation: N/A

SECTION V - FIRE-FIGHTING MEASURES

Use extinguisher suitable to cause of fire and its surroundings.
Suitable extinguishing media: N/A
Special firefighting measures: N/A
Special protection equipment: N/A
Forbidden extinguishing media: N/A
Unusual fire and explosion hazards: N/A

SECTION VI - ACCIDENTAL RELEASE MEASURES

Transfer bulk into suitable containers for disposal. Wipe residue with damp cloth and scrape up into suitable container for disposal.
Personal precautions: N/A
Reclaiming methods: N/A
Environmental precautions: N/A

SECTION VII - HANDLING AND STORAGE

Handling: No handling problems are likely to be encountered provided normal safe practices of hygiene are applied.
Storage: Store in a cool dry place.
Incompatibilities: N/A
Containers: N/A
Other precautions: N/A

SECTION VIII - EXPOSURE CONTROLS / PERSONAL PROTECTION

General measures: Wash hands thoroughly after contact with this product.
Limits: N/A
Respiratory protection: N/A
Skin protection: Overalls, rubber gloves.
Other equipment: N/A
Eye protection: Goggles/eyeshields.
Hands protection: Overalls, rubber gloves.

SECTION IX - PHYSICAL AND CHEMICAL PROPERTIES

Appearance: White stiff paste
Water solubility: Insoluble
Flash point: N/A
Odor: Odorless
Boiling point: N/A
Explosion limits: N/A
Evaporation rate: N/A

SECTION X - STABILITY AND REACTIVITY

Stability: Stable, non-reactive
Materials to avoid: N/A
Hazardous polymerization products: N/A
Conditions to avoid: N/A
Hazardous decomposition products: N/A

SECTION XI - TOXICOLOGICAL INFORMATION

Ingestion: N/A
Skin contact: N/A
Prolonged exposure: N/A
Eye contact: N/A
Inhalation: N/A

SECTION XII - ECOLOGICAL INFORMATION**SECTION XIII - DISPOSAL CONSIDERATIONS**

Dispose of in accordance with all federal, state, and local regulations.

SECTION XIV - TRANSPORT INFORMATION**SECTION XV - REGULATORY INFORMATION**

Hazard Classification Symbol: N/A
Ingredients determining the label: The information in this MSDS does not constitute an assessment of workplace risk.

SECTION XVI - OTHER INFORMATION

Handling of product exclusively by dental specialists according to instructions for use.

CAUTION: PRODUCT FOR PROFESSIONAL USE ONLY

The above information is based on presently available data and to our best knowledge for handling the product under normal conditions. Any use of this product in any way not indicated on this document or using it together with any other process/procedure will be exclusively under the user's responsibility.

**CENTRIX
INCORPORATED**

770 RIVER ROAD, SHELTON, CT 06484 USA
(203) 929-5582 (800) 235-5862 FAX (203) 929-6804

CENTRIX BV
BERGWEG 169, 3707 AC ZEIST, THE NETHERLANDS
TEL: +31 (0)30 692 70 50 FAX: +31 (0)30 693 14 21